



AB Diagnostic Systems

REF

DK.336.01.8



96

IVD

For *In vitro* Diagnostic Use

INSTRUCTIONS FOR USE

abia HBc Ab

**Enzyme immunoassay for detection of total antibodies
to Hepatitis B core antigen**

This Package Insert provides information for Professional Use of the kit.

The kit contains sufficient reagents for 96 assays (one breakable plate) including controls; the kit is intended for manual testing with a possibility of fractional (one strip) use of the kit or use of the kit on open type automated analyzer for enzyme immunoassay.

I. INTENDED USE

The abia HBc Ab kit is an enzyme immunoassay for the detection of total antibodies to Hepatitis B core antigen in human serum or plasma.

This kit is for diagnostic use by a trained laboratory professional and will not be sold to the general public. All the reagents are for professional *in vitro* diagnostic use only.

The results of this or any other diagnostic assay should be used and interpreted only in the context of the overall clinical picture.

II. INTRODUCTION

Detection of antibodies to Hepatitis B core antigen (HBc), in combination with antibodies against HBs is the most prevalent method of acute, chronic or past HBV infection diagnostics confirmation.

Antibodies against Hepatitis B core antigen appear at the onset of acute HBV infection, but may also be detectable at different stages of seroconversion and infection, indicating chronic HBV infection.

III. PRINCIPLE OF THE TEST

abia HBc Ab is a competitive solid-phase enzyme assay based on microwells coated with recombinant Hepatitis B core antigens. The conjugate contains HRP-labeled human anti-HBc antibodies.

Serum or plasma samples and HRP-labeled human anti-HBc antibodies (conjugate) are added to the wells. The specific anti-HBc antibodies present in a sample and the HRP-labeled human anti-HBc antibodies (conjugate) compete on forming stable complexes with the HBc antigens immobilized on the wells.

The unbound components are removed by washing. After the addition of the solution containing TMB, the wells with bound conjugate develop a blue color, which is converted to yellow after the reaction has been stopped with sulfuric acid. The color intensity is inversely proportional to the concentration of antibodies to HBc antigen in the specimen and can be read at 450 nm or 450/620-680 nm filters.

IV. CONTENTS OF THE KIT abia HBc Ab

4.1. Contents of the reagent kit.

Table 1

LABEL	NATURE OF THE REAGENTS	PRESENTATION
HBcore Coated Strips	Polystyrene stripped 96-well plate (breakable wells) coated with purified recombinant Hepatitis B core antigen. Store at 2-8 °C until expiration date.	1 plate
Conjugate (concentrated 21-fold)	Antibodies against human anti-HBcore, conjugated with HRP enzyme with addition of bovine serum albumin (1.7%) and Tween-20 (0.10%). Glycerol based solution. Transparent or slightly opalescent blue liquid. Preserving agents: 0.03% ProClin, 0.0009% gentamycin sulfate. Store at 2-8 °C until expiration date in a tightly sealed vial.	0.5 ml 1 vial
Positive Control, Inactivated	Control sample, containing anti-HBcore antibodies, negative for anti-HIV-1,2, HBsAg and anti-HCV. Transparent or slightly opalescent red liquid. Preserving agent: 0.02% gentamycin sulfate, 0.02% thimerosal. Store at 2-8 °C until expiration date in a tightly sealed vial.	2.0 ml 1 vial
Negative Control, Inactivated	Control sample negative for anti-HBcore, anti-HIV-1,2, HBsAg and anti-HCV. Transparent or slightly opalescent green liquid. Preserving agent: 0.01% gentamycin sulfate, 0.01% thimerosal. Store at 2-8 °C until expiration date in a tightly sealed vial.	2.5 ml 1 vial
Washing Solution (concentrated 25-fold)	Phosphate-saline solution (pH 7.4-7.7). Contains Tween-20 (2.75%). Transparent or slightly opalescent liquid, colorless or pale yellow, sediment may form that dissolves at 35-39 °C and shaking. Store at 2-8 °C until expiration date in a tightly sealed vial.	50.0 ml 1 vial

Substrate Buffer	Citric acid (0.64%) solution, pH 4.1-4.3, containing H ₂ O ₂ (0.02%). Transparent colorless liquid. Preserving agent: 0.04% ProClin. Store at 2-8 °C until expiration date in a tightly sealed vial.	25.0 ml 1 vial
TMB (concentrated 21-fold)	Solution containing 3,3',5,5'-Tetramethylbenzidine (TMB) (0.40%) and dimethyl sulfoxide (DMSO) (84.50%). Transparent colorless liquid. Store at 2-8 °C until expiration date in a tightly sealed vial.	2.5 ml 1 vial
Stopping Reagent	Sulfuric acid solution (H ₂ SO ₄) 0.75M. Transparent colorless liquid. Store at 2-8 °C until expiration date in a tightly sealed vial.	25.0 ml 1 vial
Protective films for EIA plates		1
Polyethylene bag with a Zip-Lock		1
Disposable plastic dishes for liquid reagents		2
Disposable tips		16

V. PRECAUTIONS

The reliability of the results depends on correct implementation of the following requirements:

- The temperature in the lab should be 18-24 °C.
- Inspect the contents of the box: check the vials and labels integrity. In case of label loss or labels/vials damage, vials should be disposed and **kit cannot be used**.
- Do not use expired reagents.
- Do not mix reagents from different lots within a given test run.
- Carefully reconstitute the reagents avoiding any contamination.
- Do not carry out the test in the presence of reactive vapours (acid, alkaline, aldehyde vapours) or dust that could alter the enzyme activity of the conjugates.
- Use glassware thoroughly washed and rinsed with distilled or deionized water or preferably, disposable material.
- Do not allow the microplate to dry between the end of the washing operation and the reagent distribution.
- The enzyme reaction is very sensitive to metal ions. Consequently, do not allow any metal element to come into contact with the various conjugate or substrate solutions.
- Use a new distribution tip for each sample.

- Do not reuse protective films for EIA plates.
- Well washing is a critical step in this procedure: respect the recommended number of washing cycles and make sure that all wells are completely filled and then completely emptied. Incorrect washing may lead to inaccurate results.
- Never use the same container to distribute Conjugate and other solutions.
- Check the pipettes and other equipment for accuracy and correct operation.
- Do not change the assay's procedure.
- Use distilled or deionized water.
- Avoid exposure of the reagents to excessive heat or sunlight during storage and incubation.
- Once the assay has been started, all subsequent steps should be performed without interruption.

VI. HEALTH AND SAFETY INSTRUCTIONS

- All reagents included in the kit are intended for “*in vitro* diagnostic use”.
- Human origin material used in the preparation of Positive Control and Negative Control has been tested and found non-reactive for Hepatitis B surface antigen (HBsAg), anti-HCV and anti-HIV-1,2.
- Certain reagents contain biological material of animal origin.
- Because no known test method can offer complete assurance that infections agents are absent, handle reagents and patients samples as if capable of transmitting infections disease.
- Do not eat, drink, smoke, or apply cosmetics where immunodiagnostic materials are being handled.
- Any equipment directly in contact with specimens and reagents as well as washing solutions should be considered as contaminated products and treated as such.
- Wear lab coats and disposable gloves when handling reagents and samples and thoroughly wash your hands after handling them.
- Avoid spilling samples or solutions containing samples.
- Avoid any contact of the Substrate Buffer, the TMB and the Stopping Reagent with the skin and mucosa.
- Provide adequate ventilation.
- All materials contacted with specimens or reagents, including liquid and solid wastes, should be inactivated by validated procedures (autoclaving or chemical treatment) and disposed in accordance with applicable local law regulations.

Stopping Reagent contains 0.75M sulfuric acid.



Danger!

H314 Causes severe skin burns and eye damage.

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P303 + P361 + P353 IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower.

P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes.

Remove contact lenses, if present and easy to do. Continue rinsing.

P310 Immediately call a POISON CENTER or doctor/ physician.

VII. MATERIALS AND EQUIPMENT REQUIRED BUT NOT PROVIDED WITH THE TEST:

- Distilled or deionized water.
- Automatic or semiautomatic, adjustable or preset single-channel and multi-channel pipettes with a changeable volume for a set of liquids.
- Disposable pipette tips.
- Microplate incubator at $(37.0 \pm 1.0) ^\circ\text{C}$.
- Automatic microplate washer.
- Microplate reader equipped with 450 nm or 450 nm and 620-680 nm filters.
- Open type automated analyzer with 450 nm or 450 nm and 620-680 nm filters (for automated procedure).
- Laboratory clock.

VIII. COLLECTION AND HANDLING OF SPECIMENS

Blood samples should be collected according to the current practices. The test should be performed on undiluted serum or plasma. Separate serum from red cells or plasma from the clot as soon as possible, to avoid haemolysis. Extensive haemolysis may affect test performance. Samples with observable particulate matter should be clarified by centrifugation prior testing. Suspended fibrin particles or aggregates may yield falsely positive results. Thermo inactivation of samples is not recommended as it leads to the false results.

The samples can be stored at 2-8 °C if screening is performed within 72 hours days or they may be deep-frozen at -20 °C. Avoid repeated freeze/thaw cycles. Samples that have been frozen and defrozen more than 1 times cannot be used. Samples with expressed bacterial growing, haemolysis, hyperlipidemia must not be analyzed.

IX. PREPARATION OF THE REAGENTS

1. Ready for use reagents:

- **HBcore Coated Strips.** Each plate containing 12 strips (breakable wells) is wrapped in a sealed foil-lined bag. Open the bag and remove the plate. Select the number of strips/wells required for the assay. Place the unused strips/wells back into the foil-lined bag; reseal the foil-lined bag in Zip-Lock polyethylene bag. Do not remove dessicant.

- **Negative Control;**
- **Positive Control;**
- **Stopping Reagent.**

2. Reagents to prepare:

- **Working Washing Solution.** Thoroughly mix the content of the vial with concentrated Washing Solution (concentrated 25-fold). Dilute the required volume of concentrated Washing Solution with corresponding volume of distilled or deionized water (see Table 2). Mix the solution thoroughly.
- **Working Solution of Conjugate.** The Working Solution of Conjugate should be prepared before use. Add the required volume of Working Washing Solution in a clean container and add the corresponding volume of Conjugate concentrate thoroughly mixed (see Table 2). Mix thoroughly, avoiding foaming (intensive mixing is not allowed).
- **Substrate Mixture.** Dilute the required volume of TMB (concentrated 21-fold) with the corresponding volume of Substrate Buffer (see Table 2). Mix thoroughly.

Substrate Mixture should be colorless!

The required volume of reagents for the certain number of strips or plate represented in the Table 2.

Table 2

Reagent preparation

Number of used strips	Working Washing Solution		Working Solution of Conjugate		Substrate Mixture	
	Washing Solution (concentrated 25-fold) (ml)	Deionized or distilled water (ml)	Conjugate (concentrated 21-fold) (ml)	Working Washing Solution (ml)	TMB (concentrated 21-fold) (ml)	Substrate Buffer (ml)
1	3.0	72.0	0.025	0.5	0.05	1.0
2	6.0	144.0	0.050	1.0	0.10	2.0
3	9.0	216.0	0.075	1.5	0.15	3.0
4	12.0	288.0	0.100	2.0	0.20	4.0
5	15.0	360.0	0.125	2.5	0.25	5.0
6	18.0	432.0	0.150	3.0	0.30	6.0
7	21.0	504.0	0.175	3.5	0.35	7.0
8	24.0	576.0	0.200	4.0	0.40	8.0
9	27.0	648.0	0.225	4.5	0.45	9.0
10	30.0	720.0	0.250	5.0	0.50	10.0
11	33.0	792.0	0.275	5.5	0.55	11.0
12	40.0	960.0	0.400	8.0	0.65	13.0

X. TEST PROCEDURE

Note: Before use, allow reagents to reach room temperature (18-24 °C) for 30 min.

Step	The assay procedure
1	Do not wash the coated strips before the assay. Add 50 µl of Positive Control into 1 well, and 50 µl of Negative Control into 3 wells. Add 50 µl of undiluted samples into the rest of the wells.
2	Add 50 µl of Working Solution of Conjugate into all wells. Mix the contents of the wells by careful tapping on the edge of the plate. Cover the plate protective film.
3	Incubate for 60 min in a microplate incubator at $(37.0 \pm 1.0) ^\circ\text{C}$.
4	Aspirate the contents of the wells and wash the plate 4 times with the Working Washing Solution. To each well add not less than 380 µl of Working Washing Solution, wait for 40 seconds and remove Working Washing Solution into the container with disinfecting solution. Do not leave any liquid in the wells. It is strongly recommended that an automatic microplate washer should be used. Incomplete washing will adversely affect the assay precision.
5	Add 100 µl of Substrate Mixture to all the wells of the plate.
6	Incubate at 18-24 °C for a 30 min in a dark place.
7	Add 50 µl of Stopping Reagent into each well to stop the reaction. Read the optical density at 450 nm with 620-680 nm as reference. Reading the absorbance at 450 nm only is possible.

Scheme of the assay is represented in Annex.

Automated analyzer

For automated test procedure, it is advisable to use protocol submitted by the manufacturer. When creating the protocol independently, follow the procedure specified in section X. TEST PROCEDURE, and comply with the requirements provided in section V. PRECAUTIONS.

When preparing working solutions of reagents for the automated test procedure, dead volume of vials or containers used to place the solutions onboard should be taken into account.

Validated test protocols and dilution tables of working solutions for different models of EIA analyzers can be obtained upon request from the manufacturer (see section XV.).

XI. CALCULATION OF RESULTS

The presence or absence of antibodies to Hepatitis B core antigen is determined by comparing the absorbance measured for each sample to the calculated Cut-Off value.

Test Validation

For the test to be valid the following criteria must be met. If these criteria are not met the test should be considered invalid and should be repeated.

1. **Positive Control:** the absorbance value should not be more than 0.200.
2. **Negative Control:** the absorbance value should not be less than 1.400.

Calculate Cut-Off value as:

$$\text{Cut-Off} = (\text{average OD value of Negative Control} \times 0.5) - 0.500,$$

where **0.5** and **0.500** – are a coefficients defined by manufacturer during statistical processing.

Negative: if the OD value is $>$ Cut-Off

Positive: if the OD value is \leq Cut-Off

XII. PERFORMANCE CHARACTERISTICS OF abia HBc Ab

The performance of abia HBc Ab has been determined by samples from random blood donors, samples positive to antibodies against HBc and samples from patients with confirmed various infectious diseases.

1. Diagnostic sensitivity

Category	Number of specimens	Specificity, %
HBV positive samples	52	100% (95% CI: 93.1-100%)
Samples positive to antibodies against HBcAb	57	100% (95% CI: 93.0-100%)
Total	109	100% (95% CI: 96.6-100%)

2. Diagnostic specificity

Category	Number of specimens	Specificity, %
Unselected donors	314	100% (95% CI: 98.8-100%)
HCV infected persons	69	100% (95% CI: 94.7-100%)
Pregnant women	87	97.7% (95% CI: 92.0-99.4%)
Hospitalized patients	84	98.8% (95% CI: 93.7-99.8%)
Total	554	99.5% (95% CI: 98.4-99.8%)

XIII. LIMITS OF THE TEST

- A non-reactive test result does not exclude the possibility of exposure to Hepatitis B virus.
- Results from immunosuppressed patients should be interpreted with caution.
- Testing additional HBV markers is recommended for the final diagnosis of the infection.
- The results of this or any other diagnostic assay should be used and interpreted only in the context of the overall clinical picture.

XIV. CONDITIONS OF STORAGE AND TRANSPORTATION

- **Expiry date is indicated on the packaging.** Storage and transportation conditions for the kit, conditions and terms of storage for working solutions and unused reagents are specified in Table 3.
- Transportation should be done at specified temperature in accordance with established transportation regulations. Kits transported at improper temperature cannot be used.
- Kits stored improperly cannot be used.

Table 3

1	Storage conditions			
	Keep in a dark dry place at 2-8 °C. Freezing is prohibited.			
2	Transportation conditions			
	at 2-8 °C			
	at 9-20 °C	not more than during ten (10) days		
3	Conditions and terms of storage for working solutions			
	Keep in a dark dry place and in a chemically neutral vial or in reagent container used in open automatic EIA analyzer			
	Working Solution	Washing	at 2-8 °C	For up to 28 days
			at 18-24 °C	For up to 14 days
	Working Solution of Conjugate		at 18-24 °C	For up to 12 hours
Substrate Mixture		at 18-24 °C	For up to 10 hours	
4	Conditions and terms of storage of unused reagents after opening			
	Keep in a dark dry place at 2-8 °C.			
	HBcore Coated Strips	Place the unused strips/wells back into the bag, reseal the foil-lined package in Zip-Lock polyethylene bag. Do not remove desiccant.	For 1 month	

Conjugate, Positive Control, Negative Control, Washing Solution, Substrate Buffer, TMB, Stopping Reagent	Close the vials tightly with screw caps and stored them in the manufacturer's package.	Until the kit expiration date
--	--	-------------------------------

XV. GUARANTEE

- Manufacturer guarantees conformity of the product to the requirements of regulatory and technical documentation.
- Quality and safety of the kit is guaranteed within established shelf life.
- Please contact Manufacturer if you have any questions.

















AB Diagnostic Systems GmbH

Sportfliegerstraße 4, Berlin, 12487, Germany

Tel. +49 30 208987160, Fax: +49 30 208987199

E-Mail: info@ab-ds.de, www.ab-ds.de

XVI. EXPLANATION OF SYMBOLS

	Manufacturer		Keep dry
	Date of manufacture CCYY-MM		Storage temperature limitation
	Expiry date CCYY-MM-DD		Consult Instruction for use
	Batch code		For in vitro diagnostic use
	Catalog number		Sufficient for
	Fragile, handle with care		Symbol "corrosion"
	Keep away from sunlight	Danger!	Signal word
	Top		

Scheme of the assay

1	Add	50 µl of Positive Control into 1 well, Negative Control into 3 wells
2	Add	50 µl of samples
3	Add	50 µl of Working Solution of Conjugate into all wells
4	Incubate	60 min, (37.0 ± 1.0) °C, microplate incubator
5	Wash the plate	Working Washing Solution, not less than 380 µl, 4 times
6	Add	100 µl of Substrate Mixture
7	Incubate	30 min, 18-24 °C in a dark place
8	Add	50 µl of Stopping Reagent
9	Read the optical density	450 nm/620-680 nm or 450 nm